Instructions for use:

Remove only as many test strips as are required. Reseal the container immediately after removing the strips. Do not touch the test paper zones with your fingers. Dip test strip with all test fields into swimmingpool water and make 5 back and forth movements. Shake off excess liquid. Compare immediately test field for free chlorine (test field at the end of the STRIP) and after 30 sec. test fields for total alkalinity (middle test field) and pH with colour scale.

Note:

- 1. Low pH readings may result when total alkalinity is less than 80 ppm.
- 2. Above 240 ppm the test field for total alkalinity turns blue.
- The classification of values for free chlorine, total alkalinity and pH has been done in accordance with limiting values for the USA.

Storage:

Avoid exposing the test strips to sunlight and moisture. Keep container cool and dry (storage temperature not above +30 $^{\circ}\text{C}$).